

Performance of a new NIRP TL-dosemeter -Uncertainty and detection limit estimation

NSFS Conference 2015

Roskilde, August 25

MSc, Henrik Roed

 Sundhedsstyrelsen

National Institute of Radiation Protection

TL-dosemeter & TL-reader

NIRP TL-dosemeter



Thermo Fisher Scientific,
Harshaw 8800Plus TL-reader



Harshaw 4-element TL-card



Position 1, Neutron Sensitive

Element: TLD-600H, $^6\text{LiF: Mg, Cu, P}$
3.6 mm x 0,015"

Foil: PET 12 μm / MET PET 12 μm / PE 50 μm

Position 2, $\text{H}_p(10)$

Element: TLD-700H, $^7\text{LiF: Mg, Cu, P}$
3.6 mm x 0,015"

Badge: PP / PTFE - 1000 mgcm^{-2}

Position 3, $\text{H}_p(3)$

Element: TLD-700H, $^7\text{LiF: Mg, Cu, P}$
3.6 mm x 0,015"

Badge: PP - 300 mgcm^{-2}

Position 4, $\text{H}_p(0,07)$

Element: TLD-700H, $^7\text{LiF: Mg, Cu, P}$
3.6 mm x 0,010"

Foil: PET 12 μm / MET PET 12 μm / PE 50 μm

Overview


- Evaluation of the performance (uncertainty and detection limit) by application of the Standards:
 - ISO 62387-1 "Passive integrating dosimetry systems for environmental and personal monitoring"
 - JCGM 100:2008 "Evaluation of measurement data—Guide to the expression of uncertainty in measurement" (GUM)
 - ISO 11929 "Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation"

Exposures performed by HPE (HPA/NRPB) – ISO 4037 (γ)

Linearity (ISO 62387-1)


Type Test Results, Hp(10), Cs-137

$$R = \frac{D_M}{D_T}$$




Energy Response (ISO 62387-1)

Type Test Results, Hp(10), N-20 <-> Co-60



Directional Response (ISO 62387-1)

Type Test Results, Hp(10), Vertical & Horizontal plane



Conclusion: Response within limits!

Dose uncertainty (GUM)

Model function:


$$D_{\text{gross}} = K_{E,\alpha} \cdot K_{\text{LIN}} \cdot K_{\text{TLD}} \cdot K_{\text{RCF}} \cdot K_{\text{STD}} (X - Z)$$

$$D = D_{\text{gross}} - t \cdot \dot{H}_{\text{BG}}$$


Uncertainty budget

| Quantity | Unit | Exp. value | ci | u(Xi) | Distribution | ci u(Xi) | ci u(Xi) | (ci u(Xi))^2 |
|------------------|--------|------------|---------------------|------------------------|--------------|-----------------------------|-------------|----------------|
| X | gU | 8,31 | Yg/(X-Z) | 0,006*X | Normal | 0,006 YgX/(X-Z) | 0,000661148 | 4,37117E-07 |
| Z | gU | 0,50 | Yg/(X-Z) | 0,4*Z | Triangular | 0,4 YgZ/(X-Z) | 0,002651893 | 7,03254E-06 |
| H _{BG} | mSv/d | 0,00173 | t | 0,00027 | Normal | 0,00027 t | 0,016128 | 0,000260112 |
| t | d | 60 | H _{BG} | 0,58 | Uniform | 0,58 H _{BG} | 0,000996522 | 9,93057E-07 |
| K _{LIN} | 1 | 1,04 | Yg/K _{LIN} | 0,042*K _{LIN} | Uniform | 0,042 Yg | 0,004349589 | 1,89189E-05 |
| K _{E,α} | 1 | 1,05 | Yg/K _{E,α} | 0,13*K _{E,α} | Uniform | 0,13 Yg | 0,013463014 | 0,000181253 |
| K _{TLD} | 1 | 1 | Yg/K _{TLD} | 0,006*K _{TLD} | Normal | 0,006 Yg | 0,00062137 | 3,86101E-07 |
| K _{RCF} | 1 | 1 | Yg/K _{RCF} | 0,015*K _{RCF} | Uniform | 0,015 Yg | 0,001553425 | 2,41313E-06 |
| K _{STD} | mSv/gU | 0,0121 | Yg/K _{STD} | 0,024K _{STD} | Normal | 0,024 Yg | 0,002470331 | 6,10253E-06 |
| | | | | | | | u | 0,021855171 |
| | | | | | | | u,rel | 22% |

Sensitivity coefficients



Hp(10) = 0,1 mSv




Hp(10) = 1 mSv

Minimum Detectable Dose, MDD

ISO 11929

Decision threshold: $P(X > y^* | X = 0) = \alpha$, $y^* = k_{1-\alpha}u(0)$

Detection limit: $P(X < y^* | X = y^\#) = \beta$, $y^\# = k_{1-\alpha}u(0) + k_{1-\beta}u(y^\#)$



$u(0) \neq u(y^\#)$


MDD iterative determination, Hp(10)

ISO 11929

$$\tilde{y}_{i+1} = y^* + k_{1-\beta} u(\tilde{y}_i)$$


$$y^\# = \tilde{y}_\infty$$

$$\alpha = \beta = 0,05$$



For a 1 monthly issued dosimeter MDD should be less than 0,1 mSv

Monte Carlo simulation, Confidence interval (coverage factor)








$H_p(10) = 0,1 \text{ mSv}$

$N = 100.000$


$$U = k_{0,95}(d)u(d)$$

$$k_{0,95}(0,1\text{mSv}) = 1,88$$

Coverage factor (MC)




Performance limit vs. Confidence int.



Results from intercomparison

EURADOS Report 2015-01 /2015-02



Conclusion

- Minimum Detectable Dose, $H_p(10) < 0,1 \text{ mSv}$
 - Main Influence parameters:
 - Background radiation variation
 - Dosimeter dependence on energy & angle of radiation
- Dosimeter capable of measuring $H_p(10)$:
 - Energy [16 keV \leftrightarrow 1,3 MeV] (γ)
 - Dose: [0,1 mSv \leftrightarrow 1 Sv]
 - Angle: [-60° \leftrightarrow +60°]

Thank you, ?